

EUR Research Information Portal

Response to "Cardiac output by pulse contour analysis does not match the increase measured by rebreathing during human spaceflight"

Published in:

Journal of Applied Physiology

Publication status and date:

Published: 01/01/2018

DOI (link to publisher):

[10.1152/jappphysiol.00003.2018](https://doi.org/10.1152/jappphysiol.00003.2018)

Document Version

Publisher's PDF, also known as Version of record

Citation for the published version (APA):

Houwelingen, M., & Langewouters, G. (2018). Response to "Cardiac output by pulse contour analysis does not match the increase measured by rebreathing during human spaceflight". *Journal of Applied Physiology*, 125(1), 226-227.
<https://doi.org/10.1152/jappphysiol.00003.2018>

[Link to publication on the EUR Research Information Portal](#)

Terms and Conditions of Use

Except as permitted by the applicable copyright law, you may not reproduce or make this material available to any third party without the prior written permission from the copyright holder(s). Copyright law allows the following uses of this material without prior permission:

- you may download, save and print a copy of this material for your personal use only;
- you may share the EUR portal link to this material.

In case the material is published with an open access license (e.g. a Creative Commons (CC) license), other uses may be allowed. Please check the terms and conditions of the specific license.

Take-down policy

If you believe that this material infringes your copyright and/or any other intellectual property rights, you may request its removal by contacting us at the following email address: openaccess.library@eur.nl. Please provide us with all the relevant information, including the reasons why you believe any of your rights have been infringed. In case of a legitimate complaint, we will make the material inaccessible and/or remove it from the website.

LETTER TO THE EDITOR

Response to “Cardiac output by pulse contour analysis does not match the increase measured by rebreathing during human spaceflight”

Marc J. van Houwelingen^{1,2} and Gerard Langewouters²

¹Erasmus Medical Center, Rotterdam, The Netherlands; and ²Finapres Medical Systems, Enschede, The Netherlands

Submitted 4 January 2018; accepted in final form 25 January 2018

TO THE EDITOR: The recently published paper by Hughson et al. (4) described interesting results about an increase in cardiac output (CO) during spaceflight compared with values measured while seated before flight using different methods or inert gas rebreathing and pulse contour analysis (PCA) (1, 2, 6). However, those changes were only detected by inert gas rebreathing but not by the PCA method based on blood pressure measurements on the finger. The authors concluded that the discrepancy may be explained by the inability of the PCA method to detect CO changes caused by a volume shift to the splanchnic area.

With this letter we would like to provide some additional thoughts to explain the found discrepancies and what may have caused them.

First of all, absolute values of PCA derived CO depends on the amplitude of pressure. Obtaining PCA-derived CO without absolute accurate blood pressure would therefore be a possible source of error. Hughson et al. used a device that can calibrate a reconstructed brachial pressure (Finometer Pro) on earth. However, in space, such a calibration was not executed, as the device used there does not offer that functionality. Moreover, proper cuff application is required for measuring blood pressure adequately. On earth, cuffs must be applied by trained staff. During flight experiments onboard the International Space Station, the cuff must be applied by astronauts, and without the above mentioned calibration an error introduced by a wrong or not ideal application of the cuff could be missed.

A second source of error relates to the use of a mathematical model of aortic impedance in PCA methods, frequently being the Langewouters model (7). This model computation will likely introduce a bias in CO if not calibrated properly (6, 10). This bias should not matter when assuming a constant aortic impedance, as each astronaut is his/her own control in the reported experiments. However, the long stay in space under altered vascular loading conditions may change vascular properties as described recently by the same authors (5). Those changes may also occur in the thoracic and abdominal aorta and lead to aortic impedance changes. This is supported by the observation of reduced 24 h blood pressure in space (8, 9).

A third possible source of error is inherent to all PCA methods. These methods are based on arterial pressure only and therefore do not take into account central venous pressure. However, it was shown that central venous pressure decreases

under microgravity conditions (3, 11). As a consequence, in such conditions the estimation of SVR will actually become more accurate, contributing to an improved accuracy of PCA derived CO.

The method of rebreathing itself can increase CO (3), at which point it becomes important to obtain PCA-derived CO and rebreathing-derived CO during the same interval. Indeed, inflight measurement of CO was done simultaneously. In contrast, the preflight PCA-derived CO measurements were done in supine and seated position and at rest, whereas the preflight rebreathing technique derived CO was done in only a seated position. Posture and exercise are two determinants of CO and may therefore have undermined the conclusion by the authors.

Considering the above, the conclusion made by the authors that pulse contour analysis method-derived CO should be used with caution in space may be right, but possible limitations of the used protocol and PCA method like the lack of calibration, needed user training, and control for cuff placement, as well as blood volume shifts and atrial filling changes, could have been discussed in more detail.

The important point that needs to be made is that each method has its strengths and should be carefully selected to fit the research question at hand. Using them simultaneously may even provide new information on the human circulation in space.

DISCLOSURES

Both authors are employed by Finapres Medical Systems, the legal manufacturer of Modelflow.

AUTHOR CONTRIBUTIONS

M.J.v.H. drafted manuscript; M.J.v.H. and G.J.L. edited and revised manuscript; M.J.v.H. and G.J.L. approved final version of manuscript.

REFERENCES

1. de Wilde RB, Geerts BF, Cui J, van den Berg PC, Jansen JR. Performance of three minimally invasive cardiac output monitoring systems. *Anaesthesia* 64: 762–769, 2009. doi:10.1111/j.1365-2044.2009.05934.x.
2. de Wilde RB, Schreuder JJ, van den Berg PC, Jansen JR. An evaluation of cardiac output by five arterial pulse contour techniques during cardiac surgery. *Anaesthesia* 62: 760–768, 2007. doi:10.1111/j.1365-2044.2007.05135.x.
3. Hamilton DR, Sargsyan AE, Garcia K, Ebert DJ, Whitson PA, Feiveson AH, Alferova IV, Dulchavsky SA, Matveev VP, Bogomolov VV, Duncan JM. Cardiac and vascular responses to thigh cuffs and respiratory maneuvers on crewmembers of the International Space Station. *J Appl Physiol* (1985) 112: 454–462, 2012. doi:10.1152/jappphysiol.00557.2011.

Address for reprint requests and other correspondence: M. van Houwelingen, Erasmus University Medical Center Rotterdam, Experimental Cardiology, Thoraxcenter ee23.57c, P.O. Box 2040, NL-3000CA Rotterdam, The Netherlands (e-mail: m.vanhouwelingen@erasmusmc.nl).

4. **Hughson RLPS, Peterson SD, Yee NJ, Greaves DK.** Cardiac output by pulse contour analysis does not match the increase measured by rebreathing during human spaceflight. *J Appl Physiol* (1985) 123: 1145–1149, 2017. doi:[10.1152/jappphysiol.00651.2017](https://doi.org/10.1152/jappphysiol.00651.2017).
5. **Hughson RL, Robertson AD, Arbeille P, Shoemaker JK, Rush JW, Fraser KS, Greaves DK.** Increased postflight carotid artery stiffness and in-flight insulin resistance resulting from 6-mo spaceflight in male and female astronauts. *Am J Physiol Heart Circ Physiol* 310: H628–H638, 2016. doi:[10.1152/ajpheart.00802.2015](https://doi.org/10.1152/ajpheart.00802.2015).
6. **Jansen JR, Schreuder JJ, Mulier JP, Smith NT, Settels JJ, Wesseling KH.** A comparison of cardiac output derived from the arterial pressure wave against thermodilution in cardiac surgery patients. *Br J Anaesth* 87: 212–222, 2001. doi:[10.1093/bja/87.2.212](https://doi.org/10.1093/bja/87.2.212).
7. **Langewouters GJ, Wesseling KH, Goedhard WJ.** The static elastic properties of 45 human thoracic and 20 abdominal aortas in vitro and the parameters of a new model. *J Biomech* 17: 425–435, 1984. doi:[10.1016/0021-9290\(84\)90034-4](https://doi.org/10.1016/0021-9290(84)90034-4).
8. **Norsk P, Asmar A, Damgaard M, Christensen NJ.** Fluid shifts, vasodilatation and ambulatory blood pressure reduction during long duration spaceflight. *J Physiol* 593: 573–584, 2015. doi:[10.1113/jphysiol.2014.284869](https://doi.org/10.1113/jphysiol.2014.284869).
9. **Norsk P, Damgaard M, Petersen L, Gybel M, Pump B, Gabrielsen A, Christensen NJ.** Vasorelaxation in space. *Hypertension* 47: 69–73, 2006. doi:[10.1161/01.HYP.0000194332.98674.57](https://doi.org/10.1161/01.HYP.0000194332.98674.57).
10. **Wesseling KH, Jansen JR, Settels JJ, Schreuder JJ.** Computation of aortic flow from pressure in humans using a nonlinear, three-element model. *J Appl Physiol* (1985) 74: 2566–2573, 1993. doi:[10.1152/jappl.1993.74.5.2566](https://doi.org/10.1152/jappl.1993.74.5.2566).
11. **West JB, Prisk GK.** Chest volume and shape and intrapleural pressure in microgravity. *J Appl Physiol* (1985) 87: 1240–1241, 1999. doi:[10.1152/jappl.1999.87.3.1240](https://doi.org/10.1152/jappl.1999.87.3.1240).

